

TITLE: SuperSonic Imagine Security Bulletin

RD.REC.087

Rev. A

Page 1 sur / of 1

BlueKeep Vulnerability

Publication Date: 2019-07-03

Last Update: 2019-07-03

CVSS v3.0:

Base Score: 9.8 CRITICAL

Vector: AV:N/AC:L/PR:N/UI:N/S:U/C:H/I:H/A:H

Impact Score: 5.9

Exploitability Score: 3.9

Description

On May 16, 2019, a cybersecurity vulnerability regarding Remote Desktop Services has been announced. This vulnerability identified as CVE-2019-0708 allows remote unauthenticated attackers to run arbitrary code, conduct denial of service attacks, and potentially take control of vulnerable systems.

Product Status

SuperSonic medical devices are not affected by CVE-2019-0708

Additional Information:

SuperSonic Imagine recommends the following cyber Hygiene practices:

- Ensure that your equipment is in a physically protected and actively monitored area;
- Ensure that only secure/sanitized USB storage devices are utilized;
- Ensure that your equipment is protected against network access by unsupervised systems (typically provided by mechanisms such as firewalls and VPNs); and
- Ensure your data has been backed up and stored according to your institution policy.

CISA Alert (AA19-168A) on Microsoft Operating Systems BlueKeep Vulnerability is <u>available</u> <u>here</u>.

For further information on product's security please visit our product security page: https://supersonicimagine.com/security

Les informations contenues dans ce document sont confidentielles et sont la propriété de SuperSonic Imagine (SSI). SSI met en place ce document dans l'attente qu'il soit gardé strictment confidentiel et ne doit pas être utilisé pour tout autre but qu celui qui lui est définit. Aucune portion de ce document ne doit circuler, être citée ou reproduite pour une distribution extérieure à la société sans autorisation préalable écrite de SSI.